






























# Adulterant Color Chart

Adulterant (specimen validity) strips must be read at 1 minute as pad color may change.

Test & read time	Abnormal (low)	Normal	Abnormal (high)
<b>Creatinine</b> (CR) 45 seconds	 Negative	 20  50  100  200 mg/dl	 >15mg/dl  Positive
<b>Nitrite</b> (NI) 45 seconds		 0  0.1-0.2  0.5-5.0	 >15mg/dl  Positive
<b>Glutaraldehyde</b> (GL) 45 seconds		 Negative	 Positive
<b>pH</b> (pH) Immediate	 2  3	 4  5  7  9	 ≥10
<b>Specific Gravity</b> (SG) 45 seconds	 1.000	 1.005  1.015  1.025	 ≥1.030
<b>Oxidant/Bleach</b> (OX) 30 seconds		 Negative	  Positive 